

## APPENDIX B: B DAILY CALIBRATION ANALYTICAL BALANCE

**Balance: Sartorius TE64 Analytical Balance**

YEAR: \_\_\_\_\_

[illegible]

**Document Corrective Action for Unacceptable Results on QC Deviation Form**

Calibration Weight	Accuracy $\pm$	Range
0.1 mg	0.02 mg	0.08 - 0.12
10.0 mg	0.025 mg	9.985 - 10.025
200 mg	0.06 mg	199.94-200.06
1 g	0.1 mg	.999-1.10
10 g	0.2 mg	9.8-10.2

Reviewed by/Date : \_\_\_\_\_

Reviewed by/Date : \_\_\_\_\_

Reviewed by/Date : \_\_\_\_\_

Reviewed by/Date : \_\_\_\_\_

Reviewed by/Date : \_\_\_\_\_

Reviewed by/Date : \_\_\_\_\_